



INTRODUCTION

Trulife's Seattle Select Stance Flexion Knee offers transfemoral amputees dynamic features in a lightweight package. Designed for the moderate to active amputee (K3 or K4 activity levels), Trulife's polycentric pneumatic 5-bar offers significant advantages for the amputee including: pneumatic swing phase control, adjustable stance flexion, spring knee extension assist, and adjustable flexion control and extension control. The fitting process is simplified with a built in distal tube adapter.

| Product Code | Description | Weight Limit | Proximal Attachment |
|--------------|---|-----------------|---------------------|
| SSK615 | Select Stance Flexion Knee | 100 kg (220 lb) | Pyramid |
| SSK615-THR | Select Stance Flexion Knee with Threaded Adapter | 125 kg (275 lb) | Threaded |
| SSK615-CAP | Replacement Knee Cap for SSK615 | NA | NA |

LIMITATIONS

- The SSK615 Seattle Select Flexion Knee requires a clearance of 40 mm • 1.57" from knee center to mounting surface of the knee unit. The knee reaches full flexion at 160°.
- The SSK615-THR Threaded Seattle Select Flexion Knee requires a clearance of 23 mm • 0.91" from knee center to mounting surface of the knee unit. The knee reaches full flexion at 145°.

INSTALLATION AND USE

Recommended installation and use procedures must be followed for maximum safety and service life.

Warning: Never modify the Seattle Select Stance Flexion Knee. It will void the warranty and can cause failure. Use only bolts and screws supplied by Trulife. Use of unapproved fasteners will void the warranty and can cause failure.

Proximal Socket Attachment

- The SSK615 comes with an attached proximal standard male pyramid.
- The SSK615-THR has a threaded proximal attachment point which is designed to be used either with the SCA212 4-Hole Rotatable Base, or the SCA237 3-Prong Threaded Lamination Anchor.

Pylon Attachment

1. Cut the pylon to the appropriate length. The cut must be smooth and level.
2. Remove any burrs from the end of the pylon using sand paper.
3. Fully insert the pylon into the distal knee unit.
4. Apply Loctite® 242 removable thread locking compound to the clamp bolt. Loctite® requires several hours to cure completely.
5. Using a torque wrench, tighten the Clamp Bolt (FIGURE 1) to 7.3Nm (5.4 ft-lbs or 65 in-lbs) using a 5mm hex wrench.

Warning: Never permit a patient to walk on a partially inserted or shimmed pylon. This will void the warranty and can contribute to component failure or injury.






Figure 1. SSK615 and SSK615-THR Knee adjustment locations (SSK615 Select Stance Flexion Knee shown).

Knee Alignment

When assembling the prosthesis using the Seattle Select Stance Flexion Knee, the bench alignment can be set using the following procedure.

1. Perform a Medial/Lateral (coronal plane) alignment. Set the alignment so the load line of body weight passes through the center of the knee and reaches the center of the heel.
2. Perform an Anterior/Posterior (sagittal plane) alignment. Set the alignment so the load line of body weight passes 5 mm • 0.20" forward from the anterior superior axis and the midpoint between heel and toe-break. This will be the basic alignment setting.

Note: For a highly active person, it is recommended that the alignment setting be 0 to 5 mm (0" to 0.2") forward from the anterior superior axis. For a less active person, adjust the alignment setting 5 to 10 mm (0.2" to 0.4") forward.

Caution: Improper alignment of the knee or excessive forces applied to the knee unit can result in damage and or malfunction.



Knee Adjustment

The Seattle Select Stance Flexion Knee is packaged with the extension and flexion adjustment screws at nominal (factory) setting.

Stance Phase Sensitivity Control Adjustment

1. Locate the Stance Phase Sensitivity Lock Screw on the posterior of the knee (FIGURE 1). Using a 2 mm hex wrench, loosen the lock screw. **NOTE: This screw is located vertically along the knee body**
2. Locate the Stance Phase Sensitivity Control Adjustment Screw, (FIGURE 1) on the posterior of the knee. Using a 5 mm hex wrench:
 - Rotate the adjustment screw counterclockwise to increase the overall sensitivity of the joint (decreasing stability).
 - Rotate the adjustment screw clockwise to decrease the overall sensitivity of the joint (increasing stability).
3. Re-tighten the Stance Phase Sensitivity Lock Screw.




Figure 2. Stability spring kit.

Stance Flexion Resistance Adjustment

1. Locate the two Stance Flexion Resistance Adjustment Lock Screws on the anterior of the knee (FIGURE 1). Using a 2 mm hex wrench, loosen the two lock screws.
2. Locate the Stability Adjustment screws on the anterior of the knee (FIGURE 1). Using a 5 mm hex wrench:
 - Rotate both of the adjustment screws counterclockwise to allow more stance flexion.
 - Rotate both of the adjustment screws clockwise to decrease stance flexion.
3. Re-Tighten the two Stance Flexion Resistance Adjustment Lock Screws.
 - If the knee still does not have the desired level of stability:
 - a. Locate the two Stance Flexion Resistance Adjustment Lock Screws on the anterior of the knee (FIGURE 1). Using a 2 mm hex wrench, loosen the two lock screws.
 - b. Locate the Stability Adjustment screws on the anterior of the knee (FIGURE 1). Using a 5 mm hex wrench, remove both of the Stability Adjustment screws.
 - c. Remove the Extension Spring (FIGURE 2) and replace it with one of the included Extension Springs that has the proper resistance.
 - Green — most resistance
 - Red — mid resistance
 - Blue — least resistance
 - d. Re-insert the Stability Adjustment Screws using a 2 mm hex wrench and rotate them clockwise until the desired amount of stance flexion resistance obtained.
 - e. Re-tighten the Stance Flexion Resistance Adjustment Lock Screws (FIGURE 1).

Swing Flexion and Extension Control Adjustments

1. Locate the Flexion and Extension Control Adjustment Screws on the posterior of the knee (FIGURE 1). Using a 3 mm hex wrench:
 - Rotate the Flexion Control Screws clockwise to decrease the velocity of the joint's flexion.
 - Rotate the Extension Control Screws clockwise to decrease the velocity of the joint's extension.

Warning: Failure to follow the installation and use procedures set forth above may lead to structural failure of the components subjecting the user to a risk of serious personal injury.



MAINTENANCE

- Service the product at regular intervals.
- Inspect the knee for excessive wear or visual damage during normal consultations.
- Instruct patient to discontinue use and contact their physician or prosthetist if the prosthesis starts to make noise or if they experience any change in function.
- Instruct the patient to notify their physician or prosthetist if they gain a significant amount of weight.

QUESTIONS

Contact Customer Service in the U.S. at 888-878-1238, or fax 888-878-1237

Contact Customer Service in Canada at 800-267-2812, or fax 613-392-4139

If calling from outside the U.S. or Canada, contact Customer Service at 360-697-5656, or fax 360-697-6843

Visit Trulife online at www.trulife.com

LIMITED WARRANTY

Trulife warrants that the Seattle Select Stance Flexion Knee will be free from defects in material and workmanship for two (2) years from the date of installation.

This warranty will not apply if the product has been damaged by misuse, abuse, neglect, improper care, failure to follow instructions, abnormal wear and tear, or in the event that the Seattle Select Stance Flexion Knee has been modified/repared by persons unauthorized by Trulife.

If a defect in material or workmanship is found during the warranty period, Trulife will, at Trulife's option, either repair or replace the product. If it is not possible to repair or replace the product, Trulife will be limited to refunding the purchase price.

Trulife will not be liable under any legal theory for any direct, indirect, special, incidental or consequential damages arising from the use of or inability to use this product.

The application guidelines for this Trulife product are for the use of and by certified, qualified practitioner only. Patients are not to attempt to apply or adjust the item unless expressly instructed to do so by the practitioner responsible for the prescription and/or initial fitting of the device. All patient questions should be referred to the practitioner and not to the manufacturer. The manufacturer warrants only that the enclosed product has been inspected for quality and can be effective for certain indications, but final decisions and ongoing monitoring must be made by the prosthetic professional(s) prescribing and/or fitting the device to determine its effectiveness for an individual patient. Patient compliance is an integral part of the entire protocol and must be adhered to in order to avoid potential problems and to maximize the effectiveness of the prescribed product.

As a condition of the sale of any Trulife product, this prosthesis is restricted to a "Single Patient Use Only" by the originally fitted patient in order to protect the care provider and the patient against potentially adverse consequences of infectious disease transmission, material instability in adapting to the configuration of the original user and/or decrease in effectivity. Any express or implied warranties are voided if the product is reused or fitted to another patient. Additionally, a license of right to use under any relevant patents pertaining to the product is terminated with the cessation of use by the original patient. As with all Trulife products, this product must be prescribed and applied by a qualified practitioner to determine it meets the needs of the particular patient and accomplishes the desired results.

